


# Sample Date/Time Report

<b>Source Instrument</b>	TUS_Recorder	<b>Start Time</b>	16:00	<b>Stop Time</b>	19:00
<b>Source Group</b>	Logging Group	<b>Start Date</b>	28-Nov-07	<b>Stop Date</b>	28-Nov-07
<b>Data Available From</b>	28/11/2007 15:49:29	<b>Message Types</b>		<b>Duration H:M:S</b>	3:00:00
<b>To</b>	28/11/2007 19:11:34	<b>No. of Messages</b>	0		
<b>Number of Samples</b>	15 (20 Maximum)				
<b>Report generated</b>	5/27/2009 12:00				

Sample Date/Time	Controller PV °C	Controller WSP °C	Math 40 °C	Messages
28/11/2007 16:00:00	125.00	125.00	125.00	
28/11/2007 16:12:51	125.00	125.00	125.00	
28/11/2007 16:25:43	125.33	125.00	125.33	
28/11/2007 16:38:34	199.85	200.00	199.85	
28/11/2007 16:51:26	200.67	200.00	200.67	
28/11/2007 17:04:17	199.94	200.00	199.94	
28/11/2007 17:17:09	400.04	400.00	400.04	
28/11/2007 17:30:00	399.30	400.00	399.30	
28/11/2007 17:42:51	399.79	400.00	399.79	
28/11/2007 17:55:43	600.18	600.00	600.18	
28/11/2007 18:08:34	600.25	600.00	600.25	
28/11/2007 18:21:26	600.09	600.00	600.09	
28/11/2007 18:34:17	800.24	800.00	800.24	
28/11/2007 18:47:09	799.80	800.00	799.80	
28/11/2007 19:00:00	799.41	800.00	799.41	


**Validation Information for this report:-**

<b>Data Source</b>	C:\Program Files\Eurotherm\Review\tus_demo.erv
<b>Instrument MAC</b>	00:0A:8D:00:29:8B
<b>Uhistory Version</b>	1.2.1003
<b>Excel Uhistory Version</b>	1.1.1000
<b>Operating System</b>	Windows (32-bit) NT 5.01
<b>Excel Version</b>	12.0

Generated for you by